



ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337
Fax : +6 03-5122 7191
E-mail : enquiry@isolab.com.my
Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Streptococcus Selective (Streptose) Agar

Catalog No. : ISO-2276

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Amber, slight bluish, slightly opalescent.
2. pH at 25°C : (pH specification range: 7.4 ± 0.2)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were tested and gave satisfactory growth and reaction performance as indicated with the following cultures:

Test Organisms	ATCC®	Inoculum (CFU'S)	Incubation	Result
<i>Streptococcus pyogenes</i>	19615	100-1,000	35±2°C, aerobic, 18 - 24 hrs,	Satisfactory growth.
<i>Streptococcus faecalis</i>	29212	100-1,000	35±2°C, aerobic, 18 - 24 hrs,	Satisfactory growth.
<i>Staphylococcus aureus</i>	25923	100-1,000	35±2°C, aerobic, 18 - 24 hrs,	None - poor.
<i>Escherichia coli</i>	25922	100-1,000	35±2°C, aerobic, 18 - 24 hrs,	None - poor.

A satisfactory growth corresponds with moderate to heavy growth of the test organisms.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager