

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Columbia CNA Agar with 5% Horse Blood / MacConkey No 3 Agar (Two Compartments)

Catalog No. : ISO-S1005 Manufacture Date :

Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Dark Red no Browning (COLCNAHB) & Reddish opalescent (Mac3)

2. pH at 25°C : (pH specification range: 7.3 ± 0.2) Columbia CNA Horse Blood Agar

pH at 25°C : (pH specification range: 7.1 ± 0.2) MacConkey 3 Agar

3. Sterility : Passed inspection after 5 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
				Columbia CNA Horse Blood Agar
Streptococcus pyogenes	19615	100-1,000	35±2°C, CO ₂ , 24 hours	Satisfactory growth, beta hemolysis.
Streptococcus pneumoniae	6305	100-1,000	35±2°C, CO ₂ , 24 hours	Satisfactory growth, alpha hemolysis.
Staphylococcus aureus	25923	100-1,000	35±2°C, Aerobic, 24 hours	Satisfactory growth.
Proteus mirabilis	12453	100-1,000	35±2°C, Aerobic, 24 hours	Partial to complete inhibition
1				MacConkey 3 Agar
Escherichia coli	25922	100-1,000	35±2°C, Aerobic, 24 hours	Growth, pink colonies with slight bile precipitate.
Proteus mirabilis	12453	100-1,000	35±2°C, Aerobic, 24 hours	Growth, colourless colonies, swarming inhibited (partial).
Salmonella typhimurium	14028	100-1,000	35±2°C, Aerobic, 24 hours	Growth, colourless colonies.
Enterococcus faecalis	29212	1,000-2,000	35±2°C, Aerobic, 24 hours	Inhibition (partial to complete).

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.