

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : TCBS Agar / XLD Agar (Two Compartments)

Catalog No. : ISO-S1030 Manufacture Date :
Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Green, slightly opalescent (TCBS Agar) & Red, very slightly opalescent (XLD)

2 pH at 25°C : (pH specification range: 8.6 ± 0.2) TCBS 3 pH at 25°C : (pH specification range: 7.4 ± 0.2) XLD

34 Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU'S)	Incubation	Result
				TCBS Agar
Escherichia coli	25922	$10^3 - 10^4$	35±2°C, 24 hours,	Inhibition (partial to complete).
Pseudomonas aeruginosa	27853	$10^3 - 10^4$	35±2°C, 24 hours,	Inhibition (partial to complete).
Enterococcus faecalis	29212	$10^3 - 10^4$	35±2°C, 24 hours,	Inhibition (partial to complete).
Vibrio cholerae	9459	$10^3 - 10^4$	35±2°C, 24 hours,	Satisfactory growth, colonies yellow.
Vibrio parahaemolyticus	17802	$10^3 - 10^4$	35±2°C, 24 hours,	Satisfactory growth, colonies bluish green.
				XLD Agar
Escherichia coli	25922	1,000-2,000	35±2°C, Aerobic, 24 hours	Inhibition (partial to complete), colonies yellow to yellow red with slight bile precipitate.
Salmonella typhimurium	14028	100-1,000	35±2°C, Aerobic, 24 hours	Satisfactory growth, colonies red with black centre.
Shigella flexneri	12022	100-1,000	35±2°C, Aerobic, 24 hours	Satisfactory growth, colonies red.
Enterococcus faecalis	29212	1,000-2,000	35±2°C, Aerobic, 24 hours	Inhibition (complete).

A satisfactory growth corresponds with moderate to heavy growth of the test organisms.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.