



# ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,  
Kota Kemuning, 40460 Shah Alam,  
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337  
Fax : +6 03-5122 7191  
E-mail : enquiry@isolab.com.my  
Website : http://www.isolab.com.my

Rgn No : 852288

ISO 9001:2000 Certified

## QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : CDC Anaerobe 5% Sheep Blood Agar

Catalog No. : ISO-1340

Manufacture Date :

Batch No. :

Expiry Date :

### A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Dark red, uniform mixing, no browning.
2. pH at 25°C : (pH specification range:  $7.5 \pm 0.2$ )
3. Sterility : Passed inspection after 3 days.

### B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
<i>Bacteroides fragilis</i>	25285	$10^3$ - $10^4$	35±2°C, Anaerobic, 48 hr,	Growth.
<i>Clostridium perfringens</i>	13124	$10^3$ - $10^4$	35±2°C, Anaerobic, 48 hr,	Growth, beta hemolysis.
<i>Fusobacterium nucleatum</i>	25586	$10^3$ - $10^4$	35±2°C, Anaerobic, 48 hr,	Growth.
<i>Peptostreptococcus anaerobius</i>	27337	$10^3$ - $10^4$	35±2°C, Anaerobic, 48 hr,	Growth.
<i>Prevotella melaninogenica</i>	25845	$10^3$ - $10^4$	35±2°C, Anaerobic, 48 hr,	Growth.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager