

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Columbia CNA Agar with 5% Sheep Blood

Catalog No. : ISO-1500 Manufacture Date :

Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Medium red, uniform mixing, no browning.

2. pH at 25°C : (pH specification range: 7.3 ± 0.2)

3. Sterility : Passed inspection after 5 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU'S)	Incubation	Result
Streptococcus pyogenes	19615	100-1,000	35±2°C, Aerobic,	Satisfactory growth, beta hemolysis.
			24 hr	
Streptococcus pneumoniae	6305	100-1,000	35±2°C, CO ₂ , 24 hr,	Satisfactory growth, alpha hemolysis.
Staphylococcus aureus	25923	100-1,000	35±2°C,	Satisfactory growth.
			Aerobic, 24 hr	
Proteus mirabilis	124532	100-1,000	35±2°C,	Partial to complete inhibition
			Aerobic, 24 hr,	

A satisfactory growth corresponds with moderate to heavy growth of the test organisms.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager