

**ISOLAB SDN BHD** (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337
Fax : +6 03-5122 7191
E-mail : enquiry@isolab.com.my
Website : http://www.isolab.com.my

Rgn No : 852288
ISO 9001:2000 Certified

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : GC II Agar with 1% Growth Supplement

Catalog No. : ISO-1608

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Light amber, slightly opalescent
2. pH at 25°C : (pH specification range: 7.3 ± 0.2)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Using Kirby-Bauer Method of Antimicrobial Susceptibility Testing, plates were inoculated with the test organisms and appropriate antimicrobial discs, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
<i>Neisseria gonorrhoeae</i>	43069	10 ⁸	35±2°C, 5-7% CO ₂ , 20-24 hours	Satisfactory.

A satisfactory result corresponds with typical growth of the organism following the procedure described by CLSI. Zone sizes fell within the range of acceptable zone diameter quality control limits specified by the CLSI standard M2-A6.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager