

**ISOLAB SDN BHD** (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191

E-mail : enquiry@isolab.com.my

Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Mueller Hinton II Agar with 4% NaCL

Catalog No. : ISO-1832

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Light amber, slightly opalescent.
2. pH at 25°C : (pH specification range: 7.3 ± 0.1)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
<i>Staphylococcus aureus</i>	43300	$10^5 - 10^6$	35±2°C, Aerobic, 24 hours	Growth.
<i>Staphylococcus aureus</i>	25923	$10^5 - 10^6$	35±2°C, Aerobic, 24 hours	No growth.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager