



ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337
Fax : +6 03-5122 7191
E-mail : enquiry@isolab.com.my
Website : http://www.isolab.com.my

Mgn No : 852288
ISO 9001:2000 Certified

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Mueller Hinton II Agar with 5% Lysed Sheep Blood

Catalog No. : ISO-1841

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Wine red, uniform blood mix, slightly opalescent without precipitate.
2. pH at 25°C : (pH specification range: 7.3 ± 0.2)

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated and incubated under appropriate atmospheric conditions and gave cultural responses conforming to the NCCLS standard M2-A6 and M7-A4 with the following cultures:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
<i>Streptococcus pneumoniae</i>	49619	10^7	35±2°C, 5-7% CO ₂ , 24 hours	Satisfactory.
<i>Streptococcus pyogenes</i>	49117	10^7	35±2°C, 5-7% CO ₂ , 24 hours	Satisfactory.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager