

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Regan –Lowe Charcoal agar with Cephalexin

Catalog No. : ISO-2000 Manufacture Date :
Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

Plate appearance : Charcoal Black, homogeneous, opaque
 pH at 25°C : (pH specification range: 7.4 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC ®	Inoculum (CFU'S)	Incubation	Result
Bortella pertussis	9797	100-1,000	35±2°C, Aerobic, 3 days	Good.
Bortella parapertussis	15311	100-1,000	35±2°C, Aerobic, 3 days	Good.

A satisfactory growth corresponds with moderate to heavy growth of the test organisms.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager