



ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337
Fax : +6 03-5122 7191
E-mail : enquiry@isolab.com.my
Website : http://www.isolab.com.my

Rgn No : 852288
ISO 9001:2000 Certified

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Sabouraud Dextrose Agar with Chloramphenicol

Catalog No. : ISO-2070

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Light to medium amber, slightly opalescent.
2. pH at 25°C : (pH specification range: 5.6 ± 0.2)
3. Sterility : Passed inspection after 5 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU'S)	Incubation	Result
<i>Aspergillus niger</i>	16404	100-1,000	25±2°C, Aerobic, 3 – 5 days,	Fair to heavy growth.
<i>Candida albicans</i>	60193	100-1,000	30±2°C, Aerobic, 24 hrs,	Satisfactory growth.
<i>Escherichia coli</i>	25922	1,000-2,000	30±2°C, Aerobic, 2 days,	Complete inhibition.
<i>Saccharomyces cerevisiae</i>	9763	100-1,000	30±2°C, Aerobic, 24 hrs,	Satisfactory growth.
<i>Trichophyton mentagrophytes</i>	9533	100-1,000	30±2°C, Aerobic, 2 days,	Satisfactory growth.

A satisfactory growth corresponds with moderate to heavy growth of the test organisms.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager