

## ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

## **QUALITY CONTROL CERTIFICATE**

Date : Page : 1 of 1

Product : Thayer Martin Selective Agar

Catalog No. : ISO-2330 Manufacture Date :
Batch No. : Expiry Date :

## A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Medium brown, opaque

2. pH at 25°C : (pH specification range:  $7.2 \pm 0.2$ )

3. Sterility : Passed inspection after 3 days.

## B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
Candida albicans	60193	1,000	35±2°C, CO <sub>2</sub> , 24 - 48 hours	Inhibition (partial).
Escherichia coli	25922	1,000	35±2°C, CO <sub>2</sub> , 24 - 48 hours	Marked to complete inhibition.
Neisseria gonorrhoeae	43069	100-1,000	35±2°C, CO <sub>2</sub> , 24- 48 hours	Satisfactory growth. Colonies small, opaque, grayish-white to colourless, mucoid.
Neisseria meningitidis	13090	100-1,000	35±2°C, CO <sub>2</sub> , 24 - 48 hours	Satisfactory growth. Colonies medium to large, blue-gray, mucoid.
Staphylococcus epidermidis	12228	1,000	35±2°C, CO <sub>2</sub> , 24 - 48 hours	Inhibition (partial).

A satisfactory growth corresponds with moderate to heavy growth of the test organisms.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.