

**ISOLAB SDN BHD** (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191

E-mail : enquiry@isolab.com.my

Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Amies Transport Medium with Charcoal, 5mL

Catalog No. : ISO-5044

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Plate appearance : Grey to Black charcoal colour.
2. pH at 25°C : (pH specification range: 7.2 ± 0.2)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate culture agar and atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
<i>Escherichia coli</i>	25922	10^8	35±2°C, 16-18 hours	Satisfactory.
<i>Pseudomonas aeruginosa</i>	27853	10^8	35±2°C, 16-18 hours	Satisfactory.
<i>Staphylococcus aureus</i>	25923	10^8	35±2°C, 16-18 hours	Satisfactory.
<i>Escherichia coli</i>	35218	10^8	35±2°C, 16-18 hours	Satisfactory.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager