

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Buffered Peptone Water, 5mL Manufacture Date :

Catalog No.: ISO-5586 Expiry Date:

Batch No. :

A. PHYSICAL CHARACTERISTICS

1. Medium appearance : Light amber, clear with no precipitate.

2. pH at 25°C : (pH specification range: 7.2 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU'S)	Incubation	Growth
Salmonella enteritidis	13076	10-100	35±2°C, 24 hours	Good
Salmonella typhi	19430	10 - 100	35±2°C, 24 hours	Good
Salmonella typhimurium	14028	10 – 100	35±2°C, 24 hours	Good

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager