

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Fluid Thioglycollate Medium USP, 8mL

Cat. No : ISO-5831 Manufacture Date :

Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Tube appearance : Light amber, slightly opalescent with pink upper layer.

2. pH at 25°C : (pH specification range: 7.1 ± 0.2)

Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were tested with the following cultures and gave reaction performance as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Result
Bacteroides fragilis	25285	10 – 100	30-35°C, Aerobic (tighten cap), 24-48 hr	Growth.
Staphylococcus aureus	25923	10-100	30-35°C, Aerobic (tighten cap), 24-48 hr	Growth.
USP Growth Promotion Test				
Bacillus subtilis	6633	10 - 100	30-35°C, Aerobic, 7 days,	Growth.
Bacteroides vulgaus	8482	10 – 100	30-35°C, Aerobic, 7 days,	Growth.
Candida albicans	10231	10 - 100	20-25°C, Aerobic, 7 days,	Growth.
Clostridium sporogenes	11437	10 - 100	30-35°C, Aerobic, 7 days,	Growth.
Pseudomonas aeruginosa	9027	10 - 100	30-35°C, Aerobic, 7 days,	Growth.
Staphylococcus aureus	6538	10 - 100	30-35°C, Aerobic, 7 days,	Growth.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.