

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Lim Broth (Todd Hewitt with CNA) for Group B Streptococcus, 5ml

Catalog No. : ISO-5940 Manufacture Date :
Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Broth appearance : Light amber, clear, no precipitate.

2. pH at 25°C : (pH specification range: 7.8 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Incubation	Growth
Escherichia coli	25922	35±2°C,	
		Aerobic, 24	Inhibition (partial to complete)
		hours	
Streptococcus agalactiae	12386	35±2°C,	
		Aerobic, 24	Growth(moderate to heavy)
		hours	

Representative samples of t were tested for turbidity compared to uninoculated broth.

The pH stated was obtained at the time of testing. Turbid LIM Broth was subcultured into TSA 5 % Sheep Blood agar. Group B Streptococcus properties were observed to be beta or non hemolytic with negative catalyse test and Positive CAMP test

Results reported were obtained at time of release.

QC Manager