



ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337
Fax : +6 03-5122 7191
E-mail : enquiry@isolab.com.my
Website : http://www.isolab.com.my

Rgn No : 852288
ISO 9001:2000 Certified

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Lim Broth (Todd Hewitt with CNA) for Group B Streptococcus, 5ml

Catalog No. : ISO-5940

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Broth appearance : Light amber, clear, no precipitate.
2. pH at 25°C : (pH specification range: 7.8 ± 0.2)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Incubation	Growth
<i>Escherichia coli</i>	25922	35±2°C, Aerobic, 24 hours	Inhibition (partial to complete)
<i>Streptococcus agalactiae</i>	12386	35±2°C, Aerobic, 24 hours	Growth(moderate to heavy)

Representative samples of t were tested for turbidity compared to uninoculated broth.

The pH stated was obtained at the time of testing. Turbid LIM Broth was subcultured into TSA 5 % Sheep Blood agar. Group B Streptococcus properties were observed to be beta or non hemolytic with negative catalyse test and Positive CAMP test

Results reported were obtained at time of release.

QC Manager