

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Lauryl Sulfate Broth with Durham Tube (Double Strength), 10ml

Catalog No. : ISO-5931 Manufacture Date :

Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Solution appearance : Tan to yellow, clear to slightly hazy.

2. pH at 25°C : (pH specification range: 6.8 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were diluted with 10mL of sterile purified water, then inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Recovery	Gas Production
Enterobacter aerogenes	13048	30 - 300	35±2°C, 24 hours	Good	Weak +
Enterococcus faecalis	29212	$10^4 - 10^5$	35±2°C, 24 hours	Inhibited	
Escherichia coli	25922	30 – 300	35±2°C, 24 hours	Good	+
Salmonella typhimurium	14028	30 - 300	35±2°C, 24 hours	Good	-
Staphylococcus aureus	25923	300 – 1,000	35±2°C, 24 hours	Marked to complete inhibition	-

⁺ Gas Production is defined as the presence of gas in the inverted Durham Tube with a corresponding effervescence produced when the tube is gently shaken.

A satisfactory growth corresponds with moderate to heavy growth of the test organisms.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.