

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Nutrient Agar Slants

A. PHYSICAL CHARACTERISTICS

Tube appearance : Light amber, slight hazy, no precipitate.
pH at 25°C : (pH specification range: 6.8 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. <u>BIOLOGICAL PERFORMANCE</u>

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU'S)	Incubation	Growth
Pseudomonas aeruginosa	27853	100-1,000	35±2°C, Aerobic, 18 – 48 hours	Good.
Staphylococcus aureus	25923	100-1,000	35±2°C, Aerobic, 18 – 48 hours	Good.
Shigella flexneri	12022	100-1,000	35±2°C, Aerobic, 18 – 48 hours	Good.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager