



QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Nutrient Broth, 5mL

Catalog No. : ISO-6212

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Tube appearance : Light amber, clear with no precipitate.
2. pH at 25°C : (pH specification range: 6.9 ± 0.2)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Growth
<i>Escherichia coli</i>	25922	100 – 1,000	35±2°C, Aerobic, 18-24hr,	Good.
<i>Staphylococcus aureus</i>	25923	100 – 1,000	35±2°C, Aerobic, 18-24hr,	Good.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at times of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager