

**ISOLAB SDN BHD** (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,  
Kota Kemuning, 40460 Shah Alam,  
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337  
Fax : +6 03-5122 7191  
E-mail : enquiry@isolab.com.my  
Website : http://www.isolab.com.my

Rgn No : 852288  
ISO 9001:2000 Certified

## QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : OF Basal Medium, 5mL

Catalog No. : ISO-6220

Manufacture Date :

Batch No. :

Expiry Date :

### A. PHYSICAL CHARACTERISTICS

1. Tube appearance : Green, very slightly opalescent, with no precipitate
2. pH at 25°C : (pH specification range: 6.8 ± 0.1 )
3. Sterility : Passed inspection after 3 days.

### B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Incubation	Reaction	
			Uncovered	Covered with Oil
<i>Acinetobacter baumannii</i>	19606	35±2°C, Aerobic, 24-48hr,	-	-
<i>Pseudomonas aeruginosa</i>	10145	35±-2°C, Aerobic, 24-48hr,	-	-
<i>Shigella sonnei</i>	9290	35±2°C, Aerobic, 24-48hr,	-	-

K = Alkaline reaction, green medium

A = Acid reaction, yellow medium

G = Gas production

- = Unchanged

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager