

**ISOLAB SDN BHD** (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,  
Kota Kemuning, 40460 Shah Alam,  
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191

E-mail : enquiry@isolab.com.my

Website : http://www.isolab.com.my

## QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : OF Medium with Lactose, 5mL

Catalog No. : ISO-6226

Manufacture Date :

Batch No. :

Expiry Date :

### **A. PHYSICAL CHARACTERISTICS**

- Media appearance : Green, very slightly opalescent, with no precipitate.
- pH at 25°C : (pH specification range:  $6.8 \pm 0.2$ )
- Sterility : Passed inspection after 3 days.

### **B. BIOLOGICAL PERFORMANCE**

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses indicated:

| Test Organisms                                              | ATCC® | Incubation                  | Reaction  |                  |
|-------------------------------------------------------------|-------|-----------------------------|-----------|------------------|
|                                                             |       |                             | Uncovered | Covered with Oil |
| <i>Enterobacter aerogenes</i>                               | 13048 | 35±2°C, Aerobic, 24 - 48 hr | A         | A                |
| <i>Salmonella choleraesuis</i><br><i>subs. Choleraesuis</i> | 13311 | 35±2°C, Aerobic, 24 - 48 hr | K         | K                |

K= Alkaline reaction, green medium

A = Acid reaction, yellow medium

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at times of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager