



ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191

E-mail : enquiry@isolab.com.my

Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : OF Medium with Maltose, 5mL

Catalog No. : ISO-6227

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

- Media appearance : Green, very slightly opalescent, with no precipitate.
- pH at 25°C : (pH specification range: 6.8 ± 0.2)
- Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses indicated:

Test Organisms	ATCC®	Incubation	Reaction	
			Uncovered	Covered with Oil
<i>Acinetobacter baumannii</i>	19606	35±2°C, Aerobic, 24 - 48 hr	K	K
<i>Proteus vulgaris</i>	8427	35±2°C, Aerobic, 24 - 48 hr	A	A
<i>Shigella sonnei</i>	9290	35±2°C, Aerobic, 24 - 48 hr	A	A

K= Alkaline reaction, green medium

A = Acid reaction, yellow medium

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at times of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager