

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Phenol Red Broth Base with Durham Tube, 5mL

A. PHYSICAL CHARACTERISTICS

1. Media appearance: Redish Pink

2. pH at 25° C : (pH specification range: 7.4 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses indicated for dextrose disc added

Test Organisms	ATCC®	Inoculum (CFU'S)	Growth	Color	A/K	Gas
Escherichia coli	25922	Undiluted	Good	Yellow	Acid	+
Pseudomonas aeruginosa	10145	Undiluted	Good	Pink	No change	-
Enterobacter clocae	13047	Undiluted	Good	Pink	Alkaline	-

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager