

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Rapid Selective Urea Broth, 0.5ml

A. PHYSICAL CHARACTERISTICS

1. Tube appearance : Orange to reddish orange, clear.

2. pH at 25°C : (pH specification range: 6.8 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Incubation	Result
Helicobacter pylori	43504	Aerobic, 30min. @ Room Temp.	Positive (Pink Colour).
Proteus mirabilis	12453	35±2°C, Aerobic, 1 hour	Negative (No colour change).
Escherichia coli	13090	35±2°C, Aerobic, 24 hour	Negative (No colour change).
Staphylococcus aureus	25923	35±2°C, Aerobic, 24 hour	Negative (No colour change).

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager