



ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337
Fax : +6 03-5122 7191
E-mail : enquiry@isolab.com.my
Website : http://www.isolab.com.my

Rgn No : 852288
ISO 9001:2000 Certified

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Saline Normal , 2.5 ml / 5ml

Catalog No. : ISO-6362SC / ISO-6361

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Broth appearance : Colourless, clear.
2. pH at 25°C : (pH specification range: 7.2 ± 0.2)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with $10 - 10^4$ CFU's of the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses upon subculture as indicated:

Test Organisms	ATCC®	Result
<i>Escherichia coli</i>	25922	Growth
<i>Enterococcus faecalis</i>	29212	Growth
<i>Pseudomonas aeruginosa</i>	27853	Growth
<i>Staphylococcus aureus</i>	25923	Growth

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager