

**ISOLAB SDN BHD** (428427-T)

No. 5, Jalan Anggerik Mokara 31/46,
Kota Kemuning, 40460 Shah Alam,
Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337
Fax : +6 03-5122 7191
E-mail : enquiry@isolab.com.my
Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date :

Page : 1 of 1

Product : Selenite-F Broth (for Salmonella Enrichment), 10mL (Universal Bottle)

Catalog No. : ISO-6421

Manufacture Date :

Batch No. :

Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Tube appearance : Slight amber, slight hazy.
2. pH at 25°C : (pH specification range: 7.0 ± 0.2)
3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated at $35 \pm 2^\circ\text{C}$ for 18 – 24 hours and then subcultured on MacConkey II Agar and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Recovery	Colony Colour
<i>Escherichia coli</i>	25922	100-1,000	$35 \pm 2^\circ\text{C}$, Aerobic, 24 hr	Partial to complete inhibition.	Pink with bile precipitation.
<i>Salmonella typhimurium</i>	14028	100-1,000	$35 \pm 2^\circ\text{C}$, Aerobic, 24 hr	Good.	Colourless.
<i>Shigella sonnei</i>	9290	100-1,000	$35 \pm 2^\circ\text{C}$, Aerobic, 24 hr	Good.	Colourless.

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager