

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Simmons Citrate Agar Slants

Catalog No. : ISO-6480 Manufacture Date :
Batch No. : Expiry Date :

A. PHYSICAL CHARACTERISTICS

1. Media appearance: Forest green, very slightly opalescent.

2. pH at 25°C : (pH specification range: 6.9 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses indicated:

Test Organisms	ATCC®	Incubation	Growth	Colour of Medium / Slant
Escherichia coli	25922	35±2°C, Aerobic, 24 – 48 hours	None to poor	Green
Salmonella typhimurium	14028	35±2°C, Aerobic, 24 – 48 hours	Good	Blue
Enterococcus aerogenes	13048	35±2°C, Aerobic, 24 – 48 hours	Good	Blue

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release; the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager