

ISOLAB SDN BHD (428427-T)

No. 5, Jalan Anggerik Mokara 31/46, Kota Kemuning, 40460 Shah Alam, Selangor Darul Ehsan, Malaysia.

Tel : +6 03-5122 5331 / 5122 5337

Fax : +6 03-5122 7191 E-mail : enquiry@isolab.com.my Website : http://www.isolab.com.my

QUALITY CONTROL CERTIFICATE

Date : Page : 1 of 1

Product : Thioglycollate Medium w/o Indicator-135C, 5mL

A. PHYSICAL CHARACTERISTICS

1. Tube appearance : Light amber, very slightly opalescent.

2. pH at 25°C : (pH specification range: 7.0 ± 0.2)

3. Sterility : Passed inspection after 3 days.

B. BIOLOGICAL PERFORMANCE

Representative samples of this batch were inoculated with the test organisms, incubated under appropriate atmospheric conditions and gave cultural responses as indicated:

Test Organisms	ATCC®	Inoculum (CFU's)	Incubation	Growth
Bacteroides fragilis	25285	10 – 100	35±2°C, Aerobic (tighten cap), 24 – 48 hours	Good
Candida albicans	10231	10 - 100		Good
Staphylococcus aureus	25923	10 - 100		Good
Streptococcus pyogenes	19615	10 - 100		Good

Representative samples of this lot were tested for microbial load and were found to be acceptable.

The pH stated was obtained at time of release, the pH may vary within the stated range depending on age of product and type of pH meter and probe used.

Results reported were obtained at time of release.

QC Manager